




CERTIFICATE OF CONFORMITY

Issuer's Name / Producer : GLASSCO LABORATORY EQUIPMENTS PVT. LTD.
Issuer's Address / Producer : Manglai, P.O. Khudda Kalan, Ambala Cantt - 133001
Haryana, India
Object of Declaration : Sockets
Material : Borosilicate Glass
Purpose of Use : Socket are used in laboratories to quickly and easily fit leak-tight apparatus together from commonly available parts.

Specification of Sockets specified in Table-1, DIN 12249

Catalogue No.	Product Description	Socket size	Pack Qty	
001.202.02	Sockets Single	10/19	10	
001.202.03	Sockets Single	12/21	10	
001.202.04	Sockets Single	14/23	10	
001.202.05	Sockets Single	19/26	10	
001.202.06	Sockets Single	24/29	10	
001.202.07	Sockets Single	29/32	10	
001.202.08	Sockets Single	34/35	10	
001.202.09*	Sockets Single	40/38	10	
001.202.10	Sockets Single	45/40	10	
001.202.11*	Sockets Single	50/42	10	
001.202.12*	Sockets Single	55/44	10	

* Non DIN

The object of declaration described above is in conformity with the requirements of the following documents:

- DIN ISO 3585 Borosilicate Glass 3.3
- DIN 12249 Laboratory Glassware - Sockets

Material: The above are made from DIN ISO 3585:1998, heat resistant Boro 3.3 Glass that meet ASTM E-438, TYPE-1, CLASS- A Standard.

Compliance: These Socket comply in every respect characteristics, measures and accuracy are in full conformity With the provisions of the standard DIN 12249

Additional Information:

Glassco has installed the quality control system as per ISO 9001 and products delivered to the market are in full conformity with the technical Specification as per standard and with all fundamental requirements to such products.

Certificate No. IPK 0039.69-01 issued by intertek certification limited, A UKAS Management System Accredited Body



This document has been produced electronically & is valid without signature