



KAVALIER

VERIFICATION CERTIFICATE

01/25

Issuer's name: **KAVALIERRGLASS, a.s., Authorized Metrological Center K 9**
Issuer's address: **Křížová 1018/6, Prague 5**
Production plant: Sklářská 359, 285 06 Sázava, Czech Republic

Customer: AUXILAB S.L.
Instrument type: Butyrometer for cheese (3 g, 0-40 % fat)
Catalogue Nr.: 7343
Item Nr.: 1632438204020
Manufacturer: KAVALIERRGLASS, a.s.
Model/ Type: TCM 131/97 - 2713
Serial Nr: 1/25
Measuring range: 0-40 %
Permissible Volume Deviation: < 0.200 %
User:
Location of Instrument:

Verification Statement

The instrument was verified according to § 9(2) of Act No. 505/1990 Coll. on Metrology, as amended, and § 6 of Decree No. 262/2000 Coll., as amended. The instrument bears the official mark K 9 at locations specified in the Certificate of Approval (or EU Type Examination) of the instrument.

The validity of the verification is governed by Decree No. 345/2002 Coll., as amended. Verification validity expires under conditions outlined in § 7(2) of Decree No. 262/2000 Coll., as amended.

Date of Verification: 10. 07. 2025
Verified by: Marie Havlíčková

Note:

This certificate is issued upon the customer's special request as confirmation of verification of the specified measuring instrument

Sázava, 14. 07. 2025
Place and date of issue



Jana Prchalová, DiS *Prchalova!*
Head of Authorized Metrological Center

KAVALIERRGLASS, a.s., is authorized by the Office for Technical Standardization, Metrology and State Testing as an Authorized Metrological Center for Verification of Specified Measuring Instruments in the field of volumetric flasks and butyrometers. The official mark K 9 is assigned and affixed in accordance with relevant legislation.