




CERTIFICATE OF CONFORMITY

Issuer's Name / Producer : GLASSCO LABORATORY EQUIPMENTS PVT. LTD.
Issuer's Address / Producer : Manglai, P.O. Khudda Kalan, Ambala Cantt - 133001
Haryana, India
Object of Declaration : Burette
Material : Borosilicate Glass
Purpose of Use : Burette is used to deliver solution in precisely measured, variable volume. Burettes are used primarily for titration, to deliver one reactant until the precise end point of reaction reached.

Specification of burettes specified in Table-1, EN ISO 385:2005(D)

Catalogue No.	Product Description	Capacity mL	Division	Pack Qty	
112.202.01A	Burettes with PTFE needle valve stopcock in schellbatch strip printed in blue colour, batch certified.	10	0.05	1	
112.202.02A	Burettes with PTFE needle valve stopcock in schellbatch strip printed in blue colour, batch certified.	25	0.1	1	
112.202.03A	Burettes with PTFE needle valve stopcock in schellbatch strip printed in blue colour, batch certified.	50	0.1	1	
112.202.04A	Burettes with PTFE needle valve stopcock in schellbatch strip printed in blue colour, batch certified.	100	0.2	1	

The object of declaration described above is in conformity with the requirements of the following documents:

- DIN ISO 3585 Borosilicate Glass 3.3
- EN ISO 385:2005(D) Laboratory Glassware - Burettes

Material: The above are made from DIN ISO 3585:1998, heat resistant Boro 3.3 Glass that meet ASTM E-438, TYPE-1, CLASS- A Standard.

Compliance: These Burettes comply in every respect characteristics, measures and accuracy are in full conformity With the provisions of the standard DIN ISO 385.

Additional Information:

Glassco has installed the quality control system as per ISO 9001 and products delivered to the market are in full conformity with the technical Specification as per standard and with all fundamental requirements to such products.

Certificate No. IPK 0039.69-01 issued by intertek certification limited, A UKAS Management System Accredited Body



This document has been produced electronically & is valid without signature