



CERTIFICATE OF CONFORMITY

Issuer's Name / Producer : GLASSCO LABORATORY EQUIPMENTS PVT. LTD.
Issuer's Address / Producer : Manglai, P.O. Khudda Kalan, Ambala Cantt - 133004
Haryana, India
Object of Declaration : Beaker low Form
Material : Borosilicate Glass 3.3
Purpose of Use : Beakers are simple Container use for Stirring, mixing and heating
Liquids & have a small spout to Aid Pouring commonly used in
many laboratories.

Specification of Beakers specified in Table-1, ASTM E 960

Catalogue No.	Product Description	Capacity	Pack Qty	
229.235.07	Beaker low Form, Heavy wall with double Scale	250	12	
229.235.08	Beaker low Form, Heavy wall with double Scale	400	12	
229.235.09	Beaker low Form, Heavy wall with double Scale	600	6	
229.235.10	Beaker low Form, Heavy wall with double Scale	1000	6	
229.235.11*	Beaker low Form, Heavy wall with double Scale	2000	4	
229.235.14	Beaker low Form, Heavy wall with double Scale	4000	1	

*non ASTM

The object of declaration described above is in conformity with the requirements of the following documents:

- DIN ISO 3585 Borosilicate Glass 3.3
- ASTM E 960 Laboratory Glassware -Beakers

Material: The above are made from DIN ISO 3585:1998, heat resistant Boro 3.3 Glass that meet ASTM E-438, TYPE-1, CLASS- A Standard.

Compliance: These Beaker comply in every respect characteristics, measures and accuracy are in full Conformity With the provisions of the standard ASTM E 960.

Additional Information:

Glassco has installed the quality control system as per ISO 9001 and products delivered to the market are in full conformity with the technical Specification as per standard and with all fundamental requirements to such products.

Certificate No. IPK 0039.06 issued by intertek certification limited, A UKAS Management System Accredited Body



This document has been produced electronically & is valid without signature