

# CERTIFICATE

## of Conformity

EC Council Directive 2014/53/EU

Certificate No.: XK2211013065C

**Applicant** : DLAB Scientific Co., Ltd.  
YuAn Road 31, Airport Economic Core Zone, Shunyi District,  
Beijing101318, China

**Manufacturer** : DLAB Scientific Co., Ltd.  
YuAn Road 31, Airport Economic Core Zone, Shunyi District,  
Beijing101318, China

**Product** : PCR Thermocycler

**Model No.** : TC1000-G-Pro, SCI1000-G-Pro, 4044102100, 4043102100

**Brand Name** : DLAB

Applied Specifications/ Standards	Documentary Evidence	Result Of Test Report
ETSI EN 301 489-1 V2.2.3 (2019-11) ETSI EN 301 489-17 V3.2.4 (2020-09) EN IEC 61326-1:2021 EN IEC 61000-3-2: 2019+A1:2021 EN 61000-3-3:2013+A2:2019+AC:2022	XK2211013065E	Conform
ETSI EN 300 328 V2.2.2 (2019-07)	XK2211013065W	Conform
EN 61010-1	XK2211013065S	Conform
EN IEC 62311: 2020	XK2211013065MF	Conform

The certificate of conformity is based on an evaluation of a sample of the above mentioned product. Technical report is at the applicant disposal. This is to certify the tested sample that is in conformity with all provisions of above RED directive. It is only valid in connection with the test report number as listed. The certificate does not imply the assessment of the production and does not permit using the SiCT's logo without permission.



Authorized Signer: \_\_\_\_\_

Andy Wang, Manager

Date: December 05, 2022



**Shenzhen SiCT Technology Co., Ltd.**

202, Building 3, No.111 Huanguan Middle Road, Songyuanxia Community,  
Guanhu Street, Longhua District, Shenzhen, Guangdong, China

E-mail: info@sict-lab.com.cn Http:// www.sict-lab.com.cn