

CE DECLARATION OF CONFORMITY**Plum Eyewash pH Neutral 4.9% Phosphate****Eye irrigation fluid for first aid**

We, the manufacturer, of the products covered by this declaration, hereby declare, that the bottles with sterile irrigation fluid listed below, meet the requirements in the Council Directive 93/42/EEC of 14 June 1993, the Directive 2007/47/EC and its relevant transposition into national laws of the member states into which we place the device. Plum A/S is responsible for the Declaration of Conformity concerning the medical devices listed below:

Art No	Art Name	Volume
4752	Plum Eyewash pH Neutral 4.9% Phosphate	200 mL
4747	Plum Eyewash pH Neutral 4.9% Phosphate	500 mL
4746	Plum Eyewash pH Neutral 4.9% Phosphate Shower	1000 mL
4801	Plum Eyewash pH Neutral 4.9% Phosphate DUO	500 mL

With the intended use:

Plum Eyewash pH Neutral 4.9% Phosphate is to be used as a first aid product for eye irrigation involving accidents with acids and alkali.

We guarantee and declare that:

1. The device is a Class Is device according to Annex IX, rule 5.
2. The irrigation fluid is STERILE
3. The planning and manufacture procedures are conducted in conformity with the Company Quality System, in compliance with the provisions of Annex VII combined with Annex V to the Directive.

The notified body Presafe Denmark A/S (ID No. 0543) certifies Plum A/S.
The latest EC Certificate No. DGM-272 2019-12-03.

Assens, 2020-04-29

Plum A/S



Lise S Bollen
Quality Manager