Migration test compliance declaration

Noviglio, 15/06/2020

Kartell S.P.A Labware Division

DECLARES

That the following articles:

Product codes	Descriptions
2560	GRADUATED CYLINDER PP SERIGR., 10 ml
2561	GRADUATED CYLINDER PP SERIGR., 25 ml
2562	GRADUATED CYLINDER PP SERIGR., 50 ml
2563	GRADUATED CYLINDER PP SERIGR., 100 ml
2564	GRADUATED CYLINDER PP SERIGR., 250 ml
2565	GRADUATED CYLINDER PP SERIGR., 500 ml
2566	GRADUATED CYLINDER PP SERIGR., 1000 ml
2567	GRADUATED CYLINDER PP SERIGR., 2000 ml

Have been tested according to

- **Regulation EU 10/2011**

For the evaluation of overall migration, tests have been performed under standardized conditions OM2 (10 days at 40°C) with simulant B (acetic acid 3%)

Specifications on the use of the article.
Type or types of food with which it is intended to be put in contact: according to Annex III Regulation 10/2011 - Table 2 - SIMULANT B

Quality Dep Labware Division

This paper is an electronic print-out: signature is not requested