



KAVALIER

CERTIFICATE OF CONFORMITY

No. 181/KAG/2017

Issuer's name: **KAVALIERRGLASS, a.s.**
Issuer's address: Křížová 1018/6, Praha 5
office: Sklářská 359, 285 06 Sázava, Czech Republic
Object of the declaration: **Cylinders graduated, class B**
Cat No.: 1634/BB/100 ml

Material: Borosilicate glass **SIMAX**[®]
Purpose of use: laboratory volumetric glassware

The object of the declaration described above is in conformity with the requirements of the following documents:

- **ISO 3585 Borosilicate glass 3.3 - Properties.**
- **ISO 4788 Laboratory glassware - Graduated measuring cylinders.**

The producer confirms hereby that the characteristics, measures and accuracy of the products listed above are in full conformity with the provisions of the standard:

- Maximum permissible errors in capacity fulfill the values specified in Table 1/ISO 4788. These errors represent the maximum permissible error at any point on the scale.

Table 1 - ISO 4788:

Cat. No.	Nominal capacity	Max. permissible error \pm ml
1634BB	100 ml	1,0

Additional information:

The producer also declares that the products are safe when used in usual and proper way.

The producer has installed the Quality Assurance System according to ISO 9001 and thus guarantees that all products delivered to the market are in full conformity with the technical documentation and with all fundamental requirements to such products.

Certificate No. 04100940602 issued by TÜV CERT, Certification Body at TÜV NORD CERT GmbH.

The declaration is issued by the producer upon the buyer's request:

Auxilab, Spain

Sázava, 23. 10. 2017

.....
Place and date of issue


Ing. Viera Košíková

.....
quality manager